

**Refill of Medtronic Intrahepatic Pump #DAHS-NSCIIPR**

<b>Name:</b>	<b>Employee ID #:</b>
<b>Unit:</b>	<b>Title:</b>
PERFORMANCE CRITERIA - Unless otherwise specified all skills will be demonstrated in accordance with the appropriate UC Davis Health Policy and Procedure.	
These skills will be considered complete when all below performance criteria are completed and pages 1, 2 & 3 and have been scanned and emailed to: <a href="mailto:cppn@health.ucdavis.edu">cppn@health.ucdavis.edu</a>	

<b>Refill of Medtronic Intrahepatic Pump #DAHS-NSCIIPR</b>	<b>Date</b>	<b>Verifier Initials</b>
References: 1. <a href="#">UC Davis Health Policy 10007 Intrahepatic Implanted Pump Refill</a> 2. <a href="#">UC Davis Health Policy 10001 Hazardous Drugs (HD) (Chemo): Safe Handling/Preparation/Administration/Disposal of Waste/Spill Procedures</a> 3. <a href="#">Elsevier skills "Safe Handling of Hazardous Medications (Oncology) -CE</a> 4. <a href="#">Elsevier skills "Sterile Gloving-CE"</a> 5. <a href="https://www.medtronic.com/us-en/healthcare-professionals/products/neurological/drug-infusion-systems/synchromed-ii-clinician-programmer.html">https://www.medtronic.com/us-en/healthcare-professionals/products/neurological/drug-infusion-systems/synchromed-ii-clinician-programmer.html</a>		
Explain sterile procedure to patient		
Verification of provider's orders via electronic medical record (EMR), chemotherapy module (BEACON). If any dose reductions or programming changes are needed based on RN assessment, contact provider		
Perform telemetry with Synchromed Programmer to determine the volume of fluid remaining in the drug reservoir. Calculations are based on previous refill programming		
Perform hand hygiene		
Place patient in a supine position, ensuring patient comfort. Assess pump site for signs and symptoms of swelling, redness or tenderness; notify physician if present. Palpate pump to establish orientation and landmarks. Refer to physician and stop procedure if unable to palpate pump or grasp pump firmly		
If administering chemotherapy, perform independent double check, including verification of eight rights of medication administration as defined in <a href="#">UC Davis Health Policy 4055 Medication Administration</a> . If not administering chemotherapy, check syringe provided by pharmacy against physician order and verify 8 medication rights		
Demonstrate safe handling techniques and donning of appropriate Hazardous Drug PPE as described in <a href="#">UC Davis Health Policy 10001 Hazardous Drugs (HD) (Chemo): Safe Handling/Preparation/Administration/Disposal of Waste/Spill Procedures</a>		
Have a clean, clear workspace for sterile supplies. Perform hand hygiene and don surgical mask. Assemble the required supplies as listed in <a href="#">Attachment 1</a> of <a href="#">UC Davis Health Policy 10007 Intrahepatic Implanted Pump Refill</a>		
Using non-sterile gloves, prep skin with 3 sterile alcohol prep pads or sticks. Once dry, using aseptic technique, prep skin with one chlorhexidine swab stick or 3 povidone-iodine swab sticks. Let dry while opening packages and sterile gloves		
Open refill kit and sterile glove packages. Using sterile technique, drop stopcock into refill kit		
Perform hand hygiene. Don sterile gloves, additional PPE and assemble tubing set		
Place fenestrated drape, exposing pump site		
Locate and palpate the pump		
Place template over pump, aligning template edges with perimeter edges of pump. With empty syringe attached to stopcock and clamped tubing, insert provided needle through the template's center hole. Continue penetration until the needle stops at the bottom of the pump's septum. The titanium needle stop under the septum will damage the needle tip if excessive force is used. Stop procedure and refer to physician if needle stop cannot be reached with longest (2.0) needle provided in Synchromed refill kit. Consider fluoroscopy to assess pump access port or needle placement if unsure		

**Refill of Medtronic Intrahepatic Pump #DAHS-NSCIIPR**

<b>Name:</b>	<b>Employee ID #:</b>
<b>Unit:</b>	<b>Title:</b>
PERFORMANCE CRITERIA - Unless otherwise specified all skills will be demonstrated in accordance with the appropriate UC Davis Health Policy and Procedure.	
These skills will be considered complete when all below performance criteria are completed and pages 1, 2 & 3 and have been scanned and emailed to: <a href="mailto:cppn@health.ucdavis.edu">cppn@health.ucdavis.edu</a>	
Unclamp tubing and withdraw the fluid from the reservoir using gentle, negative pressure. Empty the reservoir completely (i.e., until air bubbles are present in the extension tubing). The amount withdrawn should approximately equal the previously noted reservoir volume from the current pump status readout from the programmer. Approximately 0.5 ml of fluid will remain in the extension tubing. If there is greater than 1 ml discrepancy between calculated and measured pump residual volumes, RN will consult physician. If fluid removed is chemotherapy, waste drug according to <a href="#">UC Davis Health Policy 1630 Pharmaceutical Waste Management</a> .	
Close the clamp and stopcock and remove stopcock and syringe containing residual medication. Note: The needle and tubing must remain in place	
Attach the syringe containing the prescribed fluid to the clamped extension tubing set. Verify needle placement to ensure that needle is accurately placed at the bottom of the pump	
Open the clamp and slowly inject the fluid into the reservoir in 5 ml increments. Do not force the injection. Excessive pressure caused by a full reservoir or too rapid a fill rate may cause damage to the pump or affect infusion accuracy	
Close the tubing clamp and carefully remove the needle from the pump septum	
Apply pressure to needle site with 4x4 gauze pad for a full minute or until bleeding stops	
Remove cleansing agent from skin using soap and water, if appropriate	
Ensure bleeding has stopped; apply adhesive bandage if necessary	
Dispose of all components of refill kit into appropriate waste containers	
Using Medtronic SynchroMed Programmer, perform interrogation of pump; reprogram appropriate parameters per order	
1. Use EMR After Visit Summary (AVS) to outline home cares and education needed for patient and family: <ul style="list-style-type: none"> <li>a. Purpose and use of the SynchroMed infusion pump.</li> <li>b. Possible side effects to watch for with any medication, potential problems and how to deal with them at home.</li> <li>c. Patient should concur with pump alarm date and next refill date (pump must be refilled every 14 days while patient is on active treatment. It may be possible to change to 21 days once active treatment is completed)</li> </ul>	
Documentation of the procedure should include: <ul style="list-style-type: none"> <li>a. Anticipated reservoir fluid volume calculated by the SynchroMed programmer</li> <li>b. Actual reservoir fluid aspirated from the pump</li> <li>c. Medication Administration on the Medication Administration Record</li> <li>d. Any problems with any portion of the procedure</li> </ul>	

**Refill of Medtronic Intrahepatic Pump #DAHS-NSCIIPR**

<b>Name:</b>	<b>Employee ID #:</b>
<b>Unit:</b>	<b>Title:</b>
PERFORMANCE CRITERIA - Unless otherwise specified all skills will be demonstrated in accordance with the appropriate UC Davis Health Policy and Procedure.	
These skills will be considered complete when all below performance criteria are completed and pages 1, 2 & 3 and have been scanned and emailed to: <a href="mailto:cppn@health.ucdavis.edu">cppn@health.ucdavis.edu</a>	

**PRECEPTOR SIGNATURE:**

Signature and Printed Name of Verifier (preceptor or other verified personnel) who have initialed on this form:

<b>Initials:</b>	<b>Print Name:</b>	<b>Signature:</b>

**PRECEPTEE STATEMENT AND SIGNATURE:**

I have read and understand the appropriate UC Davis Health Policies/Procedures and/or equipment operations manual, I have demonstrated the ability to perform the verified skills as noted, and I have the knowledge of the resources available to answer questions.

<b>Name:</b>	<b>Signature:</b>	<b>Date:</b>
--------------	-------------------	--------------